



त्रिभुवन विश्वविद्यालय
शिक्षण अस्पताल

तार-दुधमेड
महाराजगञ्ज
काठमाडौं, नेपाल ।

पत्र संख्या :-

मिति :

सूचना

२०८१।१२।०७

यस त्रि.वि.शिक्षण अस्पतालको ENT ओ.टी. र प्लास्टिक सर्जरी तथा बर्न विभागको ओ.टी. को लागि तपसीलमा उल्लेखित सामग्रीहरु खरिद गर्नु पर्ने भएकाले उक्त सामग्रीहरु आपूर्ति गर्न इच्छुक इजाजत प्राप्त सप्लायर्स/कम्पनी/फर्महरुले यो सूचना प्रकाशित भएको मितिले सात दिन भित्र आवश्यक सम्पूर्ण कागजात सहित शिलबन्दी दरभाउपत्र अस्पतालको सामान्य प्रशासन शाखा "क" मा पेश गर्नुहुन सूचित गरिन्छ । मूल्य विवरण र प्राविधिक विवरण फारम यसै पत्र साथ संलग्न गरिएको साथै थप जानकारीको लागि सामान्य प्रशासन शाखा 'क' मा सम्पर्क गर्न सकिनेछ ।

तपशिल :

S.N.	Particulars	Unit	Quantity
1.	Electric O.T. Table (C-Arm Compatible)	set	2 (Two)



सुमित्रा ढुङ्गेल

प्रमुख

सामान्य प्रशासन शाखा 'क'

Prakash



1. Price Schedule for Machine

1	2	3	4	5	6	7	8
Item	Description	Unit	Quantity	Unit price (Site Delivery)	Total price in figure (cols. 4 x 5)	Total price in words	Remarks
1	Electric O.T. Table (C-Arm Compatible)	Set	TWO				
				Total Amount			
				Add 13% Value Added Tax			
				Total Including VAT			

Total Price (in words)

Signature and Stamp of Bidder _____

Note: In case of discrepancy between unit price and total, the unit price shall prevail

2. Schedule of Requirements

The delivery schedule expressed as days/weeks/months stipulates hereafter a delivery date which is the date of delivery to the final destination where the Goods is required to be delivered.

No.	Description	Quantity	Place of Delivery	Delivery schedule days/weeks/months from date of Purchase Order
	Electric O.T. Table (C-Arm Compatible)	2(Two) Set	TUTH, Maharajgunj, Ktm.	Seven Days

3. Technical Specifications of C arm Compatible Electrical Operated OT Table For Plastic Surgery & ENT:

Purchaser's Specifications		Bidder's Compliance sheet
Name of the bidder's		
Manufacturer		
Brand		
Type/Model		
Country of Origin		
S.N.	Technical Specifications	
1.	Description of Function	
1.1.	Electrical Operated operating tables are tables for performing surgical procedures and it works with electrical power.	
2.	Operational Requirements	
2.1.	OT Table is required for Plastic surgery ,general surgery, Laparoscopy, Spinal surgery, Neurosurgery, ENT, Urology, Ob/Gyno surgery.	
2.2.	Electric Operating Table with complete standard accessories.	
3.	Technical Specifications	
3.1.	At least Four section table top including foot section.	
3.2.	It should make of SS-304 medical Class materials with Durable and Corrosion resistant properties.	
3.3.	The form pad should made with high density PU mattress.	
3.4.	The form should have antibacterial, antistatic, and waterproof properties.	
3.5.	Table should have manual Floor Locking system for Stability and movements .	
3.6.	Table positioning, i.e., height, lateral tilt, Trendelenburg, and reverse-Trendelenburg, height adjustment must be operated electrically.	
3.7.	It should have lifting capacity of 170kg or more.	
4	Dimensions (approx.)	
4.1	Height(adjustable): 750mm-1000mm	
4.2	Lateral tilt: Minimum $\pm 20^{\circ}$	
4.3	Trendelenburg /Reverse -Trendelenburg: $\pm 25^{\circ}/\pm 25^{\circ}$	
4.4	Leg section: $0^{\circ}-90^{\circ}$	
4.5	Back Section: $+60^{\circ}$	
4.6	Maximum width: 540 mm or more	
4.7	Length: 1900 mm or more	
5	Mechanical Function	
5.1	Head board: detachable	
5.2	Leg board: detachable	
6	Accessories, Spare Parts and Consumables included	
6.1	All standard accessories, consumables and parts required to operate the equipment should be supplied.	
6.2	Hand remote -1	
6.3	Arm support – 1 pair	

Prakash



6.4	Body strap -1	
6.5	Shoulder support – 1 pair	
6.6	Leg support for lithotomy -1 pair	
6.7	Body support -1	
6.8	Fixing clamp -1	
6.9	Mattress – 1	
6.10	Anesthetic screen -1	
6.11	Head ring clamp – 2/per table.	
5	Operational Environment	
5.1.	The product offered shall be designed to be stored and to operate normally under the conditions of the Hospital's environment. The conditions include Power Supply, Climate, Temperature, Humidity, etc.	
5.2.	Power Supply: 220 \pm 22 VAC, 50 Hz with appropriate power cable and plug.	
5.3.	The length of the power cable must be at least 3 m.	
5.4.	Ot table should have battery backup features.	
5.5.	Table Should have C-arm Compatible.	
6.	Standards, Medical Device Regulation and Safety Requirements	
6.1.	Compliance with relevant medical device regulations and standards, such as CE& ISO requirements.	
7.	User and Technical Training	
7.1.	Must provide user training (including how to use and safe handling, keep hygiene the equipment).	
7.2.	Must provide the first step technical training to troubleshooting and maintenance of the machine to the hospital Biomedical technician and engineers.	
8.	Up time Service Backup, Warranty/Guarantee	
8.1.	Should have 2 year replacement of parts & service warranty and another 1 year of free service warranty from the date of complete installation. (Delivery and installation of machine of all the item as per tender.)	
8.2.	During the warranty period supplier must ensure preventive maintenance and corrective/breakdown maintenance whenever required.	
9.	Installation and commissioning	
9.1.	The bidder must arrange for the quoted equipment with all the listed accessories and consumables to be installed on site and commissioned by certified or qualified personnel; any prerequisites for installation to be communicated to the purchaser in advance, in detail.	
10.	Documentation on site	
10.1.	User (Operating) manual in English	
10.2.	Service (Technical/ Maintenance) manual in English	