




त्रिभुवन विश्वविद्यालय
शिक्षण अस्पताल

तार-टुथमेड
महाराजगञ्ज
काठमाडौं, नेपाल ।



पत्र संख्या :-

मिति :

सूचना

२०८२/०३/०८

यस त्रि.वि.शिक्षण अस्पतालको ENT विभागको लागी १ थान Flexible NPL खरिद गर्नु पर्ने भएकाले उक्त उपकरण आपूर्ति गर्न इच्छुक इजाजत प्राप्त सप्लायर्स/कम्पनी/फर्महरुले यो सूचना प्रकाशित भएको मितिले तीन दिन भित्र आवश्यक सम्पूर्ण कागजात सहित शिलबन्दी दरभाउपत्र अस्पतालको सामान्य प्रशासन शाखा "क" मा पेश गर्नुहुन सूचित गरिन्छ । मूल्य विवरण र प्राविधिक विवरण फारम यसै पत्र साथ संलग्न गरिएको साथै थप जानकारीको लागी सामान्य प्रशासन शाखा 'क' मा सम्पर्क गर्न सकिनेछ ।

तपश्चिल :

S.N.	Particulars	Unit	Quantity
1.	Flexible NPL	Set	1 (One)

[Signature]
२०८२-०३-०८
सरस्वती गुरुङ्ग
प्रमुख
सामान्य प्रशासन 'क'

Prakash



1. Price Schedule for Machine

1	2	3	4	5	6	7	8
Item	Description	Unit	Quantity	Unit price (Site Delivery)	Total price in figure (cols. 4 x 5)	Total price in words	Remarks
1	ENT Treatment Unit.	Set	One				
Total Amount							
Add 13% Value Added Tax							
Total Including VAT							

Total Price (in words)

Signature and Stamp of Bidder _____

Note: In case of discrepancy between unit price and total, the unit price shall prevail

2. Schedule of Requirements

The delivery schedule expressed as days/weeks/months stipulates hereafter a delivery date which is the date of delivery to the final destination where the Goods is required to be delivered.

No.	Description	Quantity	Place of Delivery	Delivery schedule days/weeks/months from date of Purchase Order
	ENT Treatment Unit	1 (One) Set	TUTH, Maharajgunj, Ktm.	Seven Days

Technical Specification for Flexible NPL-1 unit

<u>Flexible NasoPharyngoLaryngoscope</u>		Bidders Offer
S. N.	Technical Specifications	
	Manufacturer:	
	Brand:	
	Type/Model:	
	Country of origin	
1	DESCRIPTION OF FUNCTION	
1.1	Portable Video Rhino laryngoscope, flexible needed in ENT Department used both in OPD and operation theater.	
2	OPERATIONAL REQUIREMENTS	
2.1	Portable Video Rhino laryngoscope, complete with all the standard accessories.	
3	SYSTEM REQUIREMENTS	
3.1	Portable Video Rhino laryngoscope, with 10 inch monitor, inbuilt LED light source, rechargeable lithium ion batteries.	
	TECHNICAL SPECIFICATIONS	
4	Adult ENT scopes	
4.1	Should have LED -based Endoscope, without any optical fiber bundle ,light source	
4.2	Shall have insertion tube diameter 3.5 mm or less.	
4.3	Shall have total working length 300 mm. (approx.)	
4.4	Field of View should be at least 100 Deg	
4.5	Shall have depth of field 3 - 50 mm. (approx.)	
4.6	Shall have angulation at least 140 degree UP, 140 degree DOWN.	
4.7	The operating handle should have at least 2 buttons for function like capture image/ Video.	
4.8	Shall have white balance facility.	
4.9	Shall have inbuilt LED light source in the video Rhino laryngoscope with 5000 hours life.	
4.10	System should support functions like recording video, photo and freeze by button on scope.	
5	Image Processor Unit:	
5.1	Bidder shall indicate brand and model information here and provide technical data document for the video processor offered	
5.2	Shall have monitor with touch screen 10 inch in size with high resolution 1280 x 720 p with digital connection port with TFT -LCD touch screen.	
5.3	Should have one touch start up for immediate use.	

Prakash



5.4	Should have three level brightness adjustment	
5.5	Should allows adjustment of color parameters	
5.6	Shall have CMOS chip on the tip providing visualization of nasal tracheal anatomy and offer high resolution still and video recording capabilities.	
5.7	Shall have feature to transfer stored still images 32 GB and videos to external memory device.	
5.8	Shall be suitable for immersion possible for disinfection.	
5.9	Shall be shock resistant ABS plastic housing.	
5.10	Shall have rechargeable Li ion battery with minimum 4 hours backup.	
5.11	All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning and lubrication materials, to be included in the offer. Bidders must specify the quantity of every item included in their offer (including items not specified above).	
5.12	<ul style="list-style-type: none"> Accessories 2 pcs brush for cleaning 1 pc air pressure leak detector+ Connector 1 pc Company made Trolley 	
6	Terms and Conditions	
6.1	Must submit ISO13485:2003/AC:2007 for Medical Devices AND	
6.2	CE (93/42 EEC Directives) and ISO Product Certificate.	
6.3	The supplier should fill the technical tender form and clearly mention the manufacturer, model no., and country of origin/Made in, else technically will be disqualified.	
6.4	The bidder must submit a valid authorization.	
6.5	During the warranty period supplier must ensure planned preventive maintenance (PPM) along with the corrective/breakdown maintenance whenever required. If equipment is down for more than 24 hours supplier must provide backup equipment until the equipment (under repair) maintains proper operation. The commitment letter of the same should be attached.	
6.6	If the technical team wants to examine physically which is not supplied in our hospital, the demonstration of the quoted model may be required during the technical evaluation. The evaluation is also based on the demonstration of the machine. If the bidder can't demonstrate the machine within the requested time, bid will be automatically disqualified.	
6.7	Should have Comprehensive 3 years warranty of complete system & 2 year additional service warranty from the date of complete installation (delivery & Installation of machine of all the items as per tender).	
6.8	All the spare parts and consumable for Intubating bronchoscope price	

Prakash



	should be submitted separately. If not bid will be disqualified.	
6.9	<p>The principle company should be responsible of fulfilling warranty/guarantee, in case local authorized agent is not able to achieve the same.</p> <p>The commitment letter of the same should be attached.</p>	
6.1 0	Onsite repair & maintenance training and operational training to the Hospital's Biomedical Engineer, Biomedical technicians and users.	
6.1 1	The machine supplied should be brand new with the date of manufacture mentioned and the country of origin should be clearly mentioned.	
6.1 2	One (Hard and soft) copy of Serve & Operating manual in English for each set should be provided at the time of installation.	
6.1 3	If the technical team wants to examine physically which is not supplied in our hospital, the demonstration of the quoted model may be required during the technical evaluation. The evaluation is also based on the demonstration of the machine. If the bidder can't demonstrate the machine within the requested time, bid will be automatically disqualified.	
6.1 4	Should have Comprehensive 3 years warranty of complete system & 2 year additional service warranty from the date of complete installation (delivery & Installation of machine of all the items as per tender).	
6.1 5	All the spare parts and consumable for Intubating bronchoscope price should be submitted separately. If not bid will be disqualified.	
6.1 6	<p>The principle company should be responsible of fulfilling warranty/guarantee, in case local authorized agent is not able to achieve the same.</p> <p>The commitment letter of the same should be attached.</p>	