



त्रिभुवन विश्वविद्यालय  
**शिक्षण अस्पताल**

तार-टुथमेड  
महाराजगञ्ज  
काठमाडौं, नेपाल।

पत्र संख्या :-



मिति : .....

२०८२।०३।०६

दोश्रो पटक प्रकाशित

यस त्रि.वि.शिक्षण अस्पतालको ENT विभाग र सुरेश बाग्ले स्मृति क्यान्सर भवनमा हेड एन्ड नेक सर्जिकल अन्कोलोजी सम्बन्धि ओ.पि.डि.संचालन गर्नको लागी १/१ थान Treatment Unit खरिद गर्नु पर्ने भएकाले उक्त सामग्रीहरु आपूर्ति गर्न इच्छुक इजाजत प्राप्त सप्लायर्स/कम्पनी/फर्महरुले यो सूचना प्रकाशित भएको मितिले तीन दिन भित्र आवश्यक सम्पूर्ण कागजात सहित शिलबन्दी दरभाउपत्र अस्पतालको सामान्य प्रशासन शाखा "क" मा पेश गर्नुहुन सूचित गरिन्छ। मूल्य विवरण र प्राविधिक विवरण फारम यसै पत्र साथ संलग्न गरिएको साथै थप जानकारीको लागी सामान्य प्रशासन शाखा 'क' मा सम्पर्क गर्न सकिनेछ।

तपशिल :

S.N.	Particulars	Unit	Quantity
1.	ENT Treatment Unit.	Set	2 (Two)

प्रा.डा.सुवास प्रसाद आचार्य  
कार्यकारी निर्देशक

*Prakash*



## 1. Price Schedule for Machine

1	2	3	4	5	6	7	8
Item	Description	Unit	Quantity	Unit price (Site Delivery)	Total price in figure (cols. 4 x 5)	Total price in words	Remarks
1	ENT Treatment Unit.	Set	TWO				
Total Amount							
Add 13% Value Added Tax							
Total Including VAT							

Total Price ..... (in words)

Signature and Stamp of Bidder \_\_\_\_\_

Note: In case of discrepancy between unit price and total, the unit price shall prevail

## 2. Schedule of Requirements

The delivery schedule expressed as days/weeks/months stipulates hereafter a delivery date which is the date of delivery to the final destination where the Goods is required to be delivered.

No.	Description	Quantity	Place of Delivery	Delivery schedule days/weeks/months from date of Purchase Order
	ENT Treatment Unit	2 (Two) Set	TUTH, Maharajgunj, Ktm.	Seven Days

## Technical Specification for ENT Treatment Unit -2 units

S N	Technical Specifications	Bidders Offer
	<b>Name of the bidder:</b>	
	<b>Manufacturer:</b>	
	<b>Country Of Origin:</b>	
	<b>Type/Brand :</b>	
	<b>Model No.</b>	
1.	<b>ENT Treatment unit should consist of.</b> <ul style="list-style-type: none"> <li>• Suction System</li> <li>• Compressed air system</li> <li>• Electric power supply for LED light handle</li> <li>• Mirror pre heater</li> </ul>	
2.	Make material should be 304 ss/ABS.	
3.	Treatment unit should consist of LED illumination light of 12v as which should have different levels of brightness adjustment.	
4.	Operating light of treatment unit should have adjustable height and angle with high intensity.	
5.	It should have spray gun with at least 30ml bottle.	
6.	It should have pre heating device for warming sinuscope, laryngoscope.	
7.	Capacity of waste bottle of treatment should be approx. 2500 ml or more.	
8.	Treatment unit should have 1 suction gun equipped with at least 3 pcs nozzle .	
9.	Air compressor must have auto-stop function.	
10.	Suction gun should start working automatically when lifted from its holder and should stop when placed back.	
11.	Treatment unit should be equipped with positive-negative regulating valve for the adjustment of the spray pressure and suction level.	
12.	<b>The configuration of treatment unit includes ;</b> <ul style="list-style-type: none"> <li>- Instrument tray,2 pcs</li> <li>- Cotton cup, 2 pcs</li> <li>- Medical bottle, 3 pcs</li> <li>- Built-in waste tank, 1 pc</li> <li>- Vacuum pump, 1 pc</li> <li>- Compressor, 1 pc</li> </ul>	
	<b>Terms and Condition</b>	
13.	Should be ISO Certified	
14.	Should provide 1 year warranty	
15.	Complete installation should be done by the bidder it self.	