



त्रिभुवन विश्वविद्यालय
शिक्षण अस्पताल
शिक्षण अस्पताल

तार-दुधमेड
महाराजगञ्ज
काठमाडौं, नेपाल।

सूचना

पत्र संख्या :-


२०८२।०६।०२

मिति :

यस त्रि.वि.शिक्षण अस्पतालको Emergency CT Scan Machine को लागी तपसीलमा उल्लेखित सामग्री खरिद गर्नु पर्ने भएकाले उक्त सामग्री आपूर्ति गर्न इच्छुक इजाजत प्राप्त सप्लायर्स/कम्पनी/फर्महरुले यो सूचना प्रकाशित भएको मितिले सात दिन भित्र आवश्यक सम्पूर्ण कागजात सहित शिलबन्दी दरभाउपत्र अस्पतालको सामान्य प्रशासन शाखा "क" मा पेश गर्नुहुन सूचित गरिन्छ। मूल्य विवरण र प्राविधिक विवरण फारम यसै पत्र साथ संलग्न गरिएको साथै थप जानकारीको लागी सामान्य प्रशासन शाखा 'क' मा सम्पर्क गर्न सकिनेछ।

तपशिल :

S.N.	Particulars	Unit	Quantity
1.	Single Head CT Scan Contrast Pressure Injector	set	1 (ONE)


सरस्वती गुरुङ्ग
प्रमुख
सामान्य प्रशासन 'क'

Prakash



1. Price Schedule for Machine

1	2	3	4	5	6	7	8
Item	Description	Unit	Quantity	Unit price (Site Delivery)	Total price in figure (cols. 4 x 5)	Total price in words	Remarks
1	Single Head CT Scan Contrast Pressure Injector.	Set	1 (One)				
Total Amount							
Add 13% Value Added Tax							
Total Including VAT							

Total Price (in words)

Signature and Stamp of Bidder _____

Note: In case of discrepancy between unit price and total, the unit price shall prevail

2. Schedule of Requirements

The delivery schedule expressed as days/weeks/months stipulates hereafter a delivery date which is the date of delivery to the final destination where the Goods is required to be delivered.

No.	Description	Quantity	Place of Delivery	Delivery schedule days/weeks/months from date of Purchase Order
	Single Head CT Scan Contrast Pressure Injector	1(One) Set	TUTH, Maharajgunj, Ktm.	Seven Days

3. Technical Specifications of Single Head CT Scan Contrast Pressure

Injector :

S.N	Purchaser's Specifications	Yes/No
	CT Injector	
	Name of the bidder	
	Manufacturer	
	Brand	
	Type/Model	
	Country of Origin	
1.	Please quote the model which includes latest state of the art equipment and meets the basic minimum specifications. The quoted model should be latest introduced.	
2.	Injector should be mains as well as rechargeable battery operated	
3.	Color touch screen (minimum size 5.5") on injector head for the easy operations	
4.	Unit should also have IR remote to start & stop the injection from the console room.	
5.	Unit should have facility for injection of Contrast and Saline- delivery of diluted contrast	
6.	Facility to store Injection history up-to at least 20 records	
7.	Injector should use 190ml of Syringes.	
8.	Injector should come with trolley mounted, 4 castors, all individually locking.	
9.	Injector head tilt approx. up (upto 80°-85°), down 60°-70° and head swivel approx. Right 130° & Left 130°	
10.	<u>Functional Requirements</u> Maximum pressure 300 psi Flow rate 0.1 to 10 ml/s Maximum volume approx. 190 ml Multi-phase protocols Up to Four phases Storage for 20 different named protocols Auto fill rate Up to 10 ml/s	
11.	Mains power operation Simultaneous operation and battery charging	
12.	<u>Consumables/Disposables required</u> 10 sets each of Syringes to be supplied along with the equipment The company should make sure to have smooth & uninterrupted supply of consumables as & when required by the institution.	
13.	Separate rate for syringe should be attached along with the bid. If the syringe rate significantly exceeds that of the syringe currently used in the existing injector, the bidder will be automatically disqualified.	
	1. Terms and Conditions	
14.	Must submit ISO13485:2003/AC:2007 for Medical Devices AND European CE approval from the notified body.	
15.	The supplier must submit the original brochures or e-copy.	
16.	The supplier should fill the technical tender form and clearly mention the manufacturer. Model no. and country of origin /Made in, else technically will be disqualified.	
17.	The bidder must submit a authorization from the manufacturer.	

Prakash



18.	Should have 3 years complete parts and service warranty and 2 year of service warranty from the date of complete installation.(delivery and installation of machine of all the item as per tender.)	
19.	All the consumable with price should be submitted separately. Bid will be disqualified, if above list is not submitted.	
20.	Must submit letter from manufacturer ensuring the availability of spare parts for at least next 8 to 10 years.	
21.	The principal company should be responsible of fulfilling warranty/guarantee. In case local authorized agent is not able to achieve the same. The commitment letter of the same should be attached.	
22.	Should provide User (Operating) manual and Service (Technical / Maintenance) manual in English.	

