



त्रिभुवन विश्वविद्यालय

शिक्षण अस्पताल

त्रिभुवन विश्वविद्यालय
शिक्षण अस्पताल

तार-टुथमेड
महाराजगञ्ज
काठमाडौं, नेपाल ।

पत्र संख्या :-

सूचना

मिति :

२०८२।०९।०४

यस त्रि.वि.शिक्षण अस्पतालको ENT O.T. को लागी आवश्यक तपसीलमा उल्लेखित सामग्रीहरु खरिद गर्नु पर्ने भएकाले उक्त सामग्रीहरु आपूर्ति गर्न इच्छुक इजाजत प्राप्त सप्लायर्स/कम्पनी/फर्महरुले यो सूचना प्रकाशित भएको मितिले सात(७) दिन भित्र आवश्यक सम्पूर्ण कागजात सहित शिलबन्दी दरभाउपत्र अस्पतालको सामान्य प्रशासन शाखा "क" मा पेश गर्नुहुन सूचित गरिन्छ । साथै थप जानकारीको लागी सामान्य प्रशासन शाखा 'क' मा सम्पर्क गर्न सकिनेछ ।

तपशिल :

S.N.	Particulars	Unit	Quantity
1.	Patient Warmer (Warming Machine)	Set	3 (Two)

Technical Specification & Condition:-

- The equipment should work on the principles of forced air/filtered air warming.
- The equipment should have 0.2 micron filter on better for air filtration
- The equipment should operate in at least three ranges of temperatures: High 43 +/-3 deg C, Med:38+/-3degC, low:32+/-3 deg C.
- The equipment should operate on 100-240VAC, 50-60HZ
- System should have same company manufactured trolley/stand with wheel for easy movement.
- The unit should be supplies with all standard accessories including warming hose, trolley/stand, power cord etc.
- The unit should have European CE and USFDA certicicate. The supplier must submit the certificate.
- The supplier muse submit the original brochure or ecopy.
- Should have 1 Years complete parts & service warranty.
- Should supply minimum 10 blankets with device.
- Must mention delivery time.

.....
सरस्वती गुरुङ्ग
प्रमुख
सामान्य प्रशासन 'क'