



त्रिभुवन विश्वविद्यालय

शिक्षण अस्पताल

तार-टुथमेड
महाराजगञ्ज
काठमाडौं, नेपाल ।

पत्र संख्या :-

मिति :



सूचना

२०८२।१२।११

यस त्रि.वि.शिक्षण अस्पतालको रक्तसञ्चार विभागको लागि Plateletes Agitator With Incubator खरिद गर्नु पर्ने भएकाले आपूर्ति गर्न इच्छुक आपूर्तिकर्ताले यो सूचना प्रकाशित भएको मितिले सात(७) दिन भित्र आवश्यक सम्पूर्ण कागजातका साथै शिलबन्दी कोटेशन यस अस्पतालको सामान्य प्रशासन शाखा "क" मा पेश गर्नुहुन सूचित गरिन्छ । मूल्य र प्राविधिक विवरण फारम पनि यसै साथ संलग्न गरिएको छ । साथै थप जानकारीको लागि सामान्य प्रशासन शाखा 'क' मा सम्पर्क गर्न सकिनेछ ।

तपशिल :

	Particulars	Unit	Req. Qty
01.	Plateletes Agitator with Incubator	Set	2 Set


.....
सरस्वती गुरुङ्ग
प्रमुख
सामान्य प्रशासन क

Prakash



1. Price Schedule for Machine

1	2	3	4	5	6	7	8
Item	Description	Unit	Quantity	Unit price (Site Delivery)	Total price in figure (cols. 4 x 5)	Total price in words	Remarks
1	Plateletes Agitator with Incubator	Set	2 (Two)				
Total Amount							
Add 13% Value Added Tax							
Total Including VAT							

Total Price (in words)

Signature and Stamp of Bidder _____

Note: In case of discrepancy between unit price and total, the unit price shall prevail

2. Schedule of Requirements

The delivery schedule expressed as days/weeks/months stipulates hereafter a delivery date which is the date of delivery to the final destination where the Goods is required to be delivered.

No.	Description	Quantity	Place of Delivery	Delivery schedule days/weeks/months from date of Purchase Order
	Plateletes Agitator with Incubator	2(Two) Set	TUTH, Maharajgunj, Ktm.	Seven Days

Prakash



Technical Specification for Platelets agitator with incubator

	Technical specification	Bidders Offer
	Name of the bidder:	
	Manufacturer:	
	Country of Origin:	
	Made in:	
	Brand:	
	Type /Model :	
	Required Technical Specification	
1.System Configuration	Platelets Agitator with incubator.	
2.Temperature controller	Microprocessor.	
3.Temperature	Preset at 22°C ;Adjustable range 20°C to 24 °C with accuracy of + - 1°C	
4.Agitator speed	Fixed or adjustable between 60 – 80 strokes per minute.	
5.Stroke amplitude	Standard 36 mm to 40 mm(approx.1.5 inches)	
6..Motion, temperature and door status Alarm	An audible alarm that triggers if the agitator motor stalls or temperature trigger at +24 °C and -20 °C	
7.Auto pause	Agitation must stop automatically when the door is opened to prevent injury to the technician.	
8.No of trays	At least 6 trays	
9..Capacity	At least 36 platelets bag (RDP), 6 Platelets bag(RDP) in a tray.	
10.Refrigeration	CFC free (Eco friendly) with heavy duty air cooled compressor.	
11.Power supply	220 – 240 V AC 50Hz fitted with appropriate plug types.	

Prakash



12.Data Management	At least 7 day temperature chart recorder.	
12..Warranty	Comprehensive warranty for 1 year from acceptance, for all the parts of instruments.	
13.Construction	Inner: stainless steel, Outer corrosion resistant material, Transparent glass door for inspection.	

