



त्रिभुवन विश्वविद्यालय  
**शिक्षण अस्पताल**  
शिक्षण अस्पताल, महाराजगञ्ज  
सूचना

तार-दुधमेड  
महाराजगञ्ज  
काठमाडौं, नेपाल।

पत्र संख्या :-

२०८३०३१०५

मिति : .....


यस त्रि.वि.शिक्षण अस्पतालको Labour Room को लागी आवश्यक १ थान Patient Warmer खरिद गर्नु पर्ने भएकाले उक्त सामग्री आपूर्ति गर्न इच्छुक इजाजत प्राप्त सप्लायर्स/कम्पनी/फर्महरुले यो सूचना प्रकाशित भएको मितिले सात दिन भित्र आवश्यक सम्पूर्ण कागजात सहित शिलबन्दी दरभाउपत्र अस्पतालको सामान्य प्रशासन शाखा "क" मा पेश गर्नुहुन सूचित गरिन्छ। मूल्य विवरण र प्राविधिक विवरण फारम यसै पत्र साथ संलग्न गरिएको साथै थप जानकारीको लागी सामान्य प्रशासन शाखा 'क' मा सम्पर्क गर्न सकिनेछ।

तपशिल :

S.N.	Particulars	Unit	Quantity
1.	Patient Warmer.	Set	1 (One)

Technical Specification & Condition:-

- The equipment should work on the principles of forced air/filtered air warming.
- The equipment should have 0.2 micron filter on better for air filtration
- The equipment should operate in at least three ranges of temperatures: High 43 +/-3 deg C, Med:38+/-3degC, low:32+/-3 deg C.
- The equipment should operate on 100-240VAC, 50-60HZ
- System should have same company manufactured trolley/stand with wheel for easy movement.
- The unit should be supplies with all standard accessories including warming hose, trolley/stand, power cord etc.
- The unit should have European CE and USFDA certiccate. The supplier must submit the certificate.
- The supplier muse submit the original brochure or ecopy.
- Should have 1 Years complete parts & service warranty.
- Should supply minimum 10 blankets with device.
- Must mention delivery time.

  
.....  
पशुपति घिमिरे  
ईन्चार्ज  
सामान्य प्रशासन शाखा 'क'