



# त्रिभुवन विश्वविद्यालय शिक्षण अस्पताल

तार-दुधमेड  
महाराजगञ्ज  
काठमाडौं, नेपाल ।

पत्र संख्या :-



मिति : .....

## सूचना

२०८३।०४।२१

यस त्रि.वि.शिक्षण अस्पतालको Ortho O.T. को लागी अती आवश्यक तपसीलमा उल्लेखित सामग्रीहरु खरिद गर्नु पर्ने भएकाले उक्त सामग्रीहरु आपूर्ति गर्न इच्छुक इजाजत प्राप्त सप्लायर्स/कम्पनी/फर्महरुले यो सूचना प्रकाशित भएको मितिले तीन (३) दिन भित्र आवश्यक सम्पूर्ण कागजात सहित शिलबन्दी दरभाउपत्र अस्पतालको सामान्य प्रशासन शाखा "क" मा पेश गर्नुहुन सूचित गरिन्छ । साथै थप जानकारीको लागी सामान्य प्रशासन शाखा 'क' मा सम्पर्क गर्न सकिनेछ ।

### तपशिल :

| S.N. | Particulars                             | Req.Qty/Unit |
|------|-----------------------------------------|--------------|
| 1.   | Fastpass Scorpion Device with 5 Needle. | 1 Set.       |
| 2.   | Suture Cutter Open, Left Notch.         | 1 Pc.        |

*Delivery Time Should be With in 10 Days*

### 1) Technical Specification of Suture Passer and Suture Cutter :-

| S.N.     | Technical Specification of Suture Passer and Suture Cutter.                           | Bidder's Offer |
|----------|---------------------------------------------------------------------------------------|----------------|
|          | <b>Manufacturer:</b>                                                                  |                |
|          | <b>Brand:</b>                                                                         |                |
|          | <b>Type/Model:</b>                                                                    |                |
|          | <b>Country of origin :</b>                                                            |                |
| <b>1</b> | <b>Suture Passer for Shoulder- 1</b>                                                  |                |
| i        | Should be reusable and autoclavable arthroscopic suture passer.                       |                |
| iii      | Should have side loading with one step suture loading.                                |                |
| iv       | Should have solid front jaw to avoid previously passed sutures.                       |                |
| v        | Should grasp tissue and pass/retrieve suture in one step.                             |                |
| vi       | Should fit through 5 and 6 mm cannula also.                                           |                |
| vii      | Should Provide Suture Passer-1 Pc. With Suture Passer Needle- 5 Pcs.                  |                |
| <b>2</b> | <b>Suture Cutter- 1</b>                                                               |                |
| i        | Should be reusable and autoclavable arthroscopic suture cutter.                       |                |
| ii       | Should be versatile for both open ended and closed ended cutting, left notch 4-4.5mm. |                |
| iii      | Should be compatible with all types of suture.                                        |                |
| iv       | Should have open-end design enabling cutting without top-loading.                     |                |
| v        | Should have left-notch facilitating intra-articular suture cutting.                   |                |

फोन नं. : ४५-१२४०४, ४५-१२५०५, ४५-१२७०७, फ्याक्स : २७७-१-४४-२२५५३, पोष्ट बक्स नं. : ३५७८  
E-mail: tuthdirector@iom.edu.np, Website: www.tuth.org.np



त्रिभुवन विश्वविद्यालय  
**शिक्षण अस्पताल**



तार-दुधमेड  
महाराजगञ्ज  
काठमाडौं, नेपाल।

पत्र संख्या :-

मिति : .....

|     |                                                                                                                                                                                                               |  |
|-----|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--|
| 3   | <b>Standard : Safety Requirements</b>                                                                                                                                                                         |  |
| i   | Must submit ISO Certificate                                                                                                                                                                                   |  |
| ii  | Must submit valid European CE (93/42 EEC Directives) and USFDA approved product certificate                                                                                                                   |  |
| 4   | <b>User Training</b>                                                                                                                                                                                          |  |
| i   | Must Provide users and Biomedical Technicians training (including how to use and maintain the equipment)                                                                                                      |  |
| 5   | <b>Warranty</b>                                                                                                                                                                                               |  |
| i   | Comprehensive Warranty for 1 year                                                                                                                                                                             |  |
| ii  | During the warranty period supplier must ensure planned preventive maintenance (PPm) along with corrective/ breakdown maintenance whenever required.                                                          |  |
| 6   | <b>GENERAL CONDITIONS AND DOCUMENTATION</b>                                                                                                                                                                   |  |
| i   | The supplier must submit the original brochure or ecopy                                                                                                                                                       |  |
| ii  | The bidder must arrange for the equipment to be installed and commissioned by certified or qualified personnel; any prerequisites for installation to be communicated to the purchaser in advance, in detail. |  |
| iii | The onsite Operational training should given by qualified technician to the users and maintenance training to the hospital's technician.                                                                      |  |
| iv  | Hard copy and soft copy of both user manual and service manual in English shall be provided at the time of installation                                                                                       |  |

.....  
सरस्वती गुरुङ्ग

प्रमुख

सामान्य प्रशासन 'क'